

C46-A2
Vol. 29 No. 8
Replaces C46-A
Vol. 21 No. 14

Blood Gas and pH Analysis and Related Measurements; Approved Guideline— Second Edition

This document provides clear definitions of the quantities in current use, and provides a single source of information on appropriate specimen collection, preanalytical variables, calibration, and quality control for blood pH and gas analysis and related measurements.

A guideline for global application developed through the Clinical and Laboratory Standards Institute consensus process.



Blood Gas and pH Analysis and Related Measurements; Approved Guideline—Second Edition

Paul D'Orazio, PhD
Sharon S. Ehrmeyer, PhD
Ellis Jacobs, PhD, DABCC, FACB
John G. Toffaletti, PhD
Jesper H. Wandrup, DrMed, Cand Scient

Abstract

This guideline is a consolidation of six CLSI documents and projects. The Area Committee on Clinical Chemistry and Toxicology concluded that CLSI's constituencies (professions, government, and industry) would be better served with the production of a single document that retains the essential information from the six original documents, while making it even more relevant and useful. It addresses blood gas, pH, and related measurements (eg, hemoglobin and hemoglobin fractions, oxygen content, hemoglobin-oxygen saturation, electrolytes, and selected metabolites) as measured in blood. It defines terminology and discusses performance characteristics as well as preanalytical variables and analytical considerations. It also addresses quality control issues.

This guideline consolidates and updates previously published CLSI/NCCLS documents C12-A—*Definitions of Quantities and Conventions Related to Blood pH and Gas Analysis; Approved Standard*; C21-A—*Performance Characteristics for Devices Measuring pO_2 and pCO_2 in Blood Samples; Approved Standard*; C25-A—*Fractional Oxyhemoglobin, Oxygen Content and Saturation, and Related Quantities in Blood: Terminology, Measurement, and Reporting; Approved Guideline*; C27-A—*Blood Gas Preanalytical Considerations: Specimen Collection and Controls; Approved Guideline*; and C32-P—*Considerations in the Simultaneous Measurement of Blood Gases, Electrolytes, and Related Analytes in Whole Blood; Proposed Guideline*; and unpublished CLSI document C33—*Practical Blood Gas and pH Quality Control*.

Sections of another CLSI/NCCLS document H11 also are included; however, H11 will remain a separate document, because its content is of interest to a broader audience.

Clinical and Laboratory Standards Institute (CLSI). *Blood Gas and pH Analysis and Related Measurements; Approved Guideline—Second Edition*. CLSI document C46-A2 (ISBN 1-56238-694-8). Clinical and Laboratory Standards Institute, 940 West Valley Road, Suite 1400, Wayne, Pennsylvania 19087-1898 USA, 2009.

The Clinical and Laboratory Standards Institute consensus process, which is the mechanism for moving a document through two or more levels of review by the health care community, is an ongoing process. Users should expect revised editions of any given document. Because rapid changes in technology may affect the procedures, methods, and protocols in a standard or guideline, users should replace outdated editions with the current editions of CLSI/NCCLS documents. Current editions are listed in the CLSI catalog and posted on our website at www.clsi.org. If your organization is not a member and would like to become one, and to request a copy of the catalog, contact us at: Telephone: 610.688.0100; Fax: 610.688.0700; E-Mail: customerservice@cls.org; Website: www.clsi.org



Contents

| | |
|---------------------------------------------------------------------------------------------------------------------------------------------------------|-----|
| Abstract..... | i |
| Committee Membership..... | iii |
| Foreword..... | vii |
| 1 Scope..... | 1 |
| 2 Introduction..... | 1 |
| 3 Standard Precautions..... | 1 |
| 4 Terminology..... | 1 |
| 4.1 Concepts and Definitions..... | 1 |
| 4.2 Abbreviations/Acronyms..... | 10 |
| 5 Preanalytical Considerations..... | 10 |
| 5.1 Patient Preparation..... | 10 |
| 5.2 Sample Device and Collection Procedures..... | 11 |
| 5.3 Transport of Specimens..... | 14 |
| 5.4 <i>In Vivo</i> Effects on Measurements..... | 15 |
| 5.5 Specimen Handling..... | 17 |
| 6 Analytical Interferences..... | 18 |
| 6.1 Interferences With Measurement of pH and Blood Gases..... | 18 |
| 6.2 Interferences With Measurement of Electrolytes (Na ⁺ , K ⁺ , Ca ⁺⁺ , Mg ⁺⁺ , and Cl ⁻)..... | 19 |
| 6.3 Interference With Measurement of Glucose and Lactate..... | 19 |
| 6.4 Interference With Measurement of Hematocrit..... | 19 |
| 6.5 Interference With Cooximetry Measurements..... | 20 |
| 6.6 Other Interferences..... | 20 |
| 7 Blood Gas Analyzer Calibration..... | 21 |
| 7.1 Principles of Calibration..... | 21 |
| 7.2 Blood Gas Analyzer Calibration..... | 21 |
| 7.3 Calibration Traceability..... | 21 |
| 7.4 Internal Electronic Barometer..... | 23 |
| 7.5 Calibration of <i>p</i> O ₂ and <i>p</i> CO ₂ in Systems for Point-of-Care Testing..... | 24 |
| 8 Blood Gas Quality Control..... | 24 |
| 8.1 Types of Quality Control..... | 24 |
| References..... | 28 |
| Appendix A. Instrument Performance Characteristics to Be Specified by the Manufacturer..... | 33 |
| Appendix B. Recommendations for Measurement and Reporting of Hemoglobin Fractions and Related Quantities..... | 36 |
| Appendix C. Measurement Technologies Used in Instruments for Analysis of Blood Gases, pH, and Related Analytes..... | 38 |
| Summary of Consensus Comments and Working Group Responses..... | 41 |

Contents (Continued)

Summary of Delegate Comments and Working Group Responses44

The Quality Management System Approach48

Related CLSI Reference Materials49

Key Words

Electrolytes, fractional hemoglobins, hemoglobin-oxygen saturation, metabolites, oxygen content, partial pressure of carbon dioxide, partial pressure of oxygen, pH

Blood Gas and pH Analysis and Related Measurements; Approved Guideline—Second Edition

1 Scope

This guideline addresses blood gas, pH, and related measurements (eg, hemoglobin and hemoglobin fractions, oxygen content, hemoglobin-oxygen saturation, electrolytes, hematocrit, glucose, and lactate) as measured in blood. The guideline is limited to devices for measurement of these quantities *in vitro*. Devices for *in vivo* monitoring and patient-attached, *ex vivo* monitors for blood gas, pH, and related measurements, although common in many respects to devices for *in vitro* measurements, are not specifically addressed.

This document defines terminology and discusses performance characteristics as well as preanalytical variables, analytical considerations, and quality control (QC) issues.

This guideline is primarily intended for laboratory technologists, respiratory therapists, critical care practitioners, and others responsible for obtaining and analyzing blood for pH, oxygen, carbon dioxide, and related measurements. It will also be useful to manufacturers and those responsible for teaching this subject to medical students, residents, and allied health personnel.

2 Introduction

Several aspects of blood pH and gas analysis are unique among clinical laboratory determinations, and, at the same time, no other test results have more immediate impact on patient care. This area of laboratory medicine also has the reputation of being somewhat confusing, partly because of the many different measured and derived quantities that have been used over the years. This document provides clear definitions of the several quantities in current use and includes information on appropriate specimen collection, preanalytical variables, and QC. There is also a section containing a list of performance characteristics pertinent to blood gas analyzers, which can be used by manufacturers to provide operational specifications in a uniform way to facilitate comparison by potential customers of different instruments.

3 Standard Precautions

Because it is often impossible to know what isolates or specimens might be infectious, all patient and laboratory specimens are treated as infectious and handled according to “standard precautions.” Standard precautions are guidelines that combine the major features of “universal precautions and body substance isolation” practices. Standard precautions cover the transmission of all infectious agents and thus are more comprehensive than universal precautions, which are intended to apply only to transmission of blood-borne pathogens. Standard and universal precaution guidelines are available from the US Centers for Disease Control and Prevention.² For specific precautions for preventing the laboratory transmission of all infectious agents from laboratory instruments and materials and for recommendations for the management of exposure to all infectious disease, refer to CLSI document M29.³

4 Terminology

4.1 Concepts and Definitions

This section contains terms and definitions in standard CLSI format integrated with related information and concepts. The formal definitions are accompanied by supplementary information necessary to

Foreword

The previous edition of this document, C46-A, was the result of the decision of the Area Committee on Clinical Chemistry and Toxicology to combine and update four approved-level documents, one proposed-level document, and one unpublished document. The intent was for this document to serve more effectively the three major constituents (professions, government, and industry) of CLSI.

C46-A consolidated the following documents:

- C12-A—*Definitions of Quantities and Conventions Related to Blood pH and Gas Analysis; Approved Standard;*
- C21-A—*Performance Characteristics for Devices Measuring pO₂ and pCO₂ in Blood Samples; Approved Standard;*
- C25-A—*Fractional Oxyhemoglobin, Oxygen Content and Saturation, and Related Quantities in Blood: Terminology, Measurement, and Reporting; Approved Guideline;*
- C27-A—*Blood Gas Preanalytical Considerations: Specimen Collection and Controls; Approved Guideline;*
- C32-P—*Considerations in the Simultaneous Measurement of Blood Gases, Electrolytes, and Related Analytes in Whole Blood; Proposed Guideline;* and
- C33—*Practical Blood Gas and pH Quality Control* (unpublished).

Sections of CLSI/NCCLS document H11¹ were also included; however, H11¹ remained a separate document, because its content includes greater detail and is of interest to a broader audience.

The current revision of the document, C46-A2, includes the following updates:

- Section 5, *Preanalytical Considerations*, was expanded to include a discussion specific to transport of specimens (see Section 5.3). Section 5.4, *In Vivo Effects on Measurements*, replaces the former section, *Patient Condition*, and was expanded.
- Section 6, *Analytical Interferences*, was expanded significantly, including references to recent literature.
- Section 7, *Blood Gas Analyzer Calibration*, was expanded significantly, including current requirements for calibration traceability.
- Section 8, *Blood Gas Quality Control*, includes newer approaches for “alternative” quality control.
- Appendix B, *Recommendations for Measurement and Reporting of Hemoglobin Fractions and Related Quantities*, was added.
- Appendix C, *Measurement Technologies Used in Instruments for Analysis of Blood Gases, pH, and Related Analytes*, was added.

The Quality Management System Approach

Clinical and Laboratory Standards Institute (CLSI) subscribes to a quality management system approach in the development of standards and guidelines, which facilitates project management; defines a document structure via a template; and provides a process to identify needed documents. The approach is based on the model presented in the most current edition of CLSI/NCCLS document HS01—*A Quality Management System Model for Health Care*. The quality management system approach applies a core set of “quality system essentials” (QSEs), basic to any organization, to all operations in any health care service’s path of workflow (ie, operational aspects that define how a particular product or service is provided). The QSEs provide the framework for delivery of any type of product or service, serving as a manager’s guide. The QSEs are:

| | | | |
|--------------------------------------------------|--------------------------------------------------------|---------------------------------------------------------------------------------------|----------------------------------------------------------------|
| Documents & Records Organization Personnel | Equipment Purchasing & Inventory Process Control | Information Management Occurrence Management Assessments—External & Internal | Process Improvement Customer Service Facilities & Safety |
|--------------------------------------------------|--------------------------------------------------------|---------------------------------------------------------------------------------------|----------------------------------------------------------------|

C46-A2 addresses the QSEs indicated by an “X.” For a description of the other documents listed in the grid, please refer to the Related CLSI Reference Materials section on the following page.

| Documents & Records | Organization | Personnel | Equipment | Purchasing & Inventory | Process Control | Information Management | Occurrence Management | Assessments—External & Internal | Process Improvement | Customer Service | Facilities & Safety |
|---------------------|--------------|-----------|-----------|------------------------|-----------------------------------------|------------------------|-----------------------|---------------------------------|---------------------|------------------|---------------------|
| GP02 | | | | | X C24 C29 EP17 | GP02 | | | | | |
| H11 | | H11 | H11 | H03 | H03 H04 H11 H18 HS03 M29 | | | | H11 | | H03 H11 M29 |

Adapted from CLSI/NCCLS document HS01—*A Quality Management System Model for Health Care*.

Path of Workflow

A path of workflow is the description of the necessary steps to deliver the particular product or service that the organization or entity provides. For example, CLSI/NCCLS document GP26—*Application of a Quality Management System Model for Laboratory Services* defines a clinical laboratory path of workflow, which consists of three sequential processes: preexamination, examination, and postexamination. All clinical laboratories follow these processes to deliver the laboratory’s services, namely quality laboratory information.

C46-A2 addresses the clinical laboratory path of workflow steps indicated by an “X.” For a description of the other documents listed in the grid, please refer to the Related CLSI Reference Materials section on the following page.

| Preexamination | | | | Examination | | | Postexamination | |
|----------------------|-------------------|------------------|---------------------------|-------------|------------------------------|----------------|---------------------------------|-------------------|
| Examination ordering | Sample collection | Sample transport | Sample receipt/processing | Examination | Results review and follow-up | Interpretation | Results reporting and archiving | Sample management |
| X | X | X | X | X | X | X | X | |
| H03 | C31 H03 H04 | C31 H03 | C31 H03 | H03 | H03 | | | |
| H11 | H11 H15 | H11 | H11 H15 H18 | H15 H18 | H15 | | | |

Adapted from CLSI/NCCLS document HS01—*A Quality Management System Model for Health Care*.

Related CLSI Reference Materials*

- C24-A3** **Statistical Quality Control for Quantitative Measurement Procedures: Principles and Definitions; Approved Guideline—Third Edition (2006).** This guideline provides definitions of analytical intervals; planning of quality control procedures; and guidance for quality control applications.
- C29-A2** **Standardization of Sodium and Potassium Ion-Selective Electrode Systems to the Flame Photometric Reference Method; Approved Standard—Second Edition (2000).** This standard contains recommendations on the expression of the results of ion-selective electrode measurement of sodium and potassium ion activities in undiluted serum, plasma, or whole-blood in clinical practice.
- C31-A2** **Ionized Calcium Determinations: Precollection Variables, Specimen Choice, Collection, and Handling; Approved Guideline—Second Edition (2001).** This document addresses preanalytical considerations, such as patient condition, specimen choice, collection, and handling—that can influence the accuracy and clinical utility of ionized calcium measurements.
- EP17-A** **Protocols for Determination of Limits of Detection and Limits of Quantitation; Approved Guideline (2004).** This document provides guidance for determining the lower limit of detection of clinical laboratory methods, for verifying claimed limits, and for the proper use and interpretation of the limits.
- GP02-A5** **Laboratory Documents: Development and Control; Approved Guideline—Fifth Edition (2006).** This document provides guidance on development, review, approval, management, and use of policy, process, and procedure documents in the medical laboratory community.
- H03-A6** **Procedures for the Collection of Diagnostic Blood Specimens by Venipuncture; Approved Standard—Sixth Edition (2007).** This document provides procedures for the collection of diagnostic specimens by venipuncture, including line draws, blood culture collection, and venipuncture in children.
- H04-A6** **Procedures and Devices for the Collection of Diagnostic Capillary Blood Specimens; Approved Standard—Sixth Edition (2008).** This document provides a technique for the collection of diagnostic capillary blood specimens, including recommendations for collection sites and specimen handling and identification. Specifications for disposable devices used to collect, process, and transfer diagnostic capillary blood specimens are also included.
- H11-A4** **Procedures for the Collection of Arterial Blood Specimens; Approved Standard—Fourth Edition (2004).** This document provides principles for collecting, handling, and transporting arterial blood specimens to assist with reducing collection hazards and ensuring the integrity of the arterial specimen.
- H15-A3** **Reference and Selected Procedures for the Quantitative Determination of Hemoglobin in Blood; Approved Standard—Third Edition (2000).** This document describes the principle, materials, and procedure for reference and standardized hemoglobin determinations. It includes specifications for secondary hemoglobinocyanide (HiCN) standards.
- H18-A3** **Procedures for the Handling and Processing of Blood Specimens; Approved Guideline—Third Edition (2004).** This document includes criteria for preparing an optimal serum or plasma sample and for the devices used to process blood specimens.
- HS03-A** **Pulse Oximetry; Approved Guideline (2005).** Pulse oximetry is a widely used device for the clinical assessment of arterial oxygenation and pulse rate. The clinical applications, quality assessment, and limitations are discussed in this guideline.
- M29-A3** **Protection of Laboratory Workers From Occupationally Acquired Infections; Approved Guideline—Third Edition (2005).** Based on US regulations, this document provides guidance on the risk of transmission of infectious agents by aerosols, droplets, blood, and body substances in a laboratory setting; specific precautions for preventing the laboratory transmission of microbial infection from laboratory instruments and materials; and recommendations for the management of exposure to infectious agents.

* Proposed-level documents are being advanced through the Clinical and Laboratory Standards Institute consensus process; therefore, readers should refer to the most current editions.