

Archived Document

This archived document is no longer being reviewed through the CLSI Consensus Document Development Process. However, this document is technically valid as of January 2017. Because of its value to the laboratory community, it is being retained in CLSI's library.



December 2000

AUTO01-A

Laboratory Automation: Specimen Container/ Specimen Carrier; Approved Standard

This document provides standards for the design and manufacture of specimen containers and carriers used for collecting and processing liquid samples, such as blood and urine, for clinical testing in laboratory automation systems.

A standard for global application developed through the Clinical and Laboratory Standards Institute consensus process.

Clinical and Laboratory Standards Institute

Setting the standard for quality in medical laboratory testing around the world.

The Clinical and Laboratory Standards Institute (CLSI) is a not-for-profit membership organization that brings together the varied perspectives and expertise of the worldwide laboratory community for the advancement of a common cause: to foster excellence in laboratory medicine by developing and implementing medical laboratory standards and guidelines that help laboratories fulfill their responsibilities with efficiency, effectiveness, and global applicability.

Consensus Process

Consensus—the substantial agreement by materially affected, competent, and interested parties—is core to the development of all CLSI documents. It does not always connote unanimous agreement, but does mean that the participants in the development of a consensus document have considered and resolved all relevant objections and accept the resulting agreement.

Commenting on Documents

CLSI documents undergo periodic evaluation and modification to keep pace with advancements in technologies, procedures, methods, and protocols affecting the laboratory or health care.

CLSI's consensus process depends on experts who volunteer to serve as contributing authors and/or as participants in the reviewing and commenting process. At the end of each comment period, the committee that developed the document is obligated to review all comments, respond in writing to all substantive comments, and revise the draft document as appropriate.

Comments on published CLSI documents are equally essential, and may be submitted by anyone, at any time, on any document. All comments are managed according to the consensus process by a committee of experts.

Appeals Process

When it is believed that an objection has not been adequately considered and responded to, the process for appeals, documented in the CLSI Standards Development Policies and Processes, is followed.

All comments and responses submitted on draft and published documents are retained on file at CLSI and are available upon request.

Get Involved—Volunteer!

Do you use CLSI documents in your workplace? Do you see room for improvement? Would you like to get involved in the revision process? Or maybe you see a need to develop a new document for an emerging technology? CLSI wants to hear from you. We are always looking for volunteers. By donating your time and talents to improve the standards that affect your own work, you will play an active role in improving public health across the globe.

For additional information on committee participation or to submit comments, contact CLSI.

Clinical and Laboratory Standards Institute
950 West Valley Road, Suite 2500
Wayne, PA 19087 USA
P: +1.610.688.0100
F: +1.610.688.0700
www.clsi.org
standard@clsi.org

ISBN 1-56238-427-9
ISSN 0273-3099

AUTO01-A
Vol. 20 No. 29
Replaces AUTO1-P
Vol. 19 No. 12

Laboratory Automation: Specimen Container/Specimen Carrier; Approved Standard

Volume 20 Number 29

Paul J. Orsulak, Ph.D., Chairholder
Keiichi Naka, Ph.D., Vice-Chairholder
Nelson S. Everhart, III
Mary Haven
L. Omar Henderson, Ph.D.
Dennis A. Lamb
Stewart Marsden

Abstract

Laboratory Automation: Specimen Container/Specimen Carrier; Approved Standard (CLSI document AUTO01-A) was developed for those engaged in the design and manufacture of specimen collection containers used for specimen handling in the healthcare and clinical laboratory environments, and for those engaged in the design and manufacture of clinical laboratory instrumentation and clinical laboratory automation systems. This document is intended to *lead* design and manufacturing toward standardized products for a wider variety of instruments and automated laboratory systems.

CLSI. *Laboratory Automation: Specimen Container/Specimen Carrier; Approved Standard*. CLSI document AUTO01-A. Clinical and Laboratory Standards Institute, 950 West Valley Road, Suite 2500, Wayne, Pennsylvania 19087 USA, 2000.

The Clinical and Laboratory Standards Institute consensus process, which is the mechanism for moving a document through two or more levels of review by the health care community, is an ongoing process. Users should expect revised editions of any given document. Because rapid changes in technology may affect the procedures, methods, and protocols in a standard or guideline, users should replace outdated editions with the current editions of CLSI documents. Current editions are listed in the CLSI catalog and posted on our website at www.clsi.org. If your organization is not a member and would like to become one, and to request a copy of the catalog, contact us at: Telephone: 610.688.0100; Fax: 610.688.0700; E-Mail: customerservice@clsi.org; Website: www.clsi.org.



Copyright ©2000 Clinical and Laboratory Standards Institute. Except as stated below, any reproduction of content from a CLSI copyrighted standard, guideline, companion product, or other material requires express written consent from CLSI. All rights reserved. Interested parties may send permission requests to permissions@clsi.org.

CLSI hereby grants permission to each individual member or purchaser to make a single reproduction of this publication for use in its laboratory procedure manual at a single site. To request permission to use this publication in any other manner, e-mail permissions@clsi.org.

Suggested Citation

CLSI. *Laboratory Automation: Specimen Container/Specimen Carrier; Approved Standard*. CLSI document AUTO01-A. Wayne, PA: Clinical and Laboratory Standards Institute; 2000.

Previous Edition:

July 1999

Reaffirmed:

March 2012

Archived:

January 2017

ISBN 1-56238-427-9

ISSN 0273-3099

Contents

Abstract.....	v
Committee Membership.....	v
Active Membership.....	vii
Matrix of NCCLS Laboratory Automation Standards.....	xiv
Preface to Laboratory Automation Standards.....	xvii
Foreword.....	xxi
1 Introduction.....	1
2 Scope.....	2
3 Definitions.....	2
4 Standard Precautions.....	12
5 Specifications for Containers.....	12
5.1 Container Body.....	13
5.2 Container Design.....	13
5.3 Closure Configurations.....	13
5.4 Tube Dimensions and Tolerances.....	13
6 Specifications for a Multiple-Specimen Container Carrier.....	14
7 Specifications for a Single-Specimen Container Carrier.....	15
References.....	16
Additional References.....	17
Appendix A. Examples of Container and Closure Configurations.....	18
Appendix B. Illustration of X, Y Pitch Orientation of the Multiple-Specimen Container Carrier.....	19
Summary of Comments and Subcommittee Responses.....	20
Summary of Delegate Voting Comments and Subcommittee Responses.....	25
Related NCCLS Publications.....	27

Matrix of NCCLS Laboratory Automation Standards

The laboratory automation standards documents, AUTO01, AUTO2, AUTO3, AUTO4, and AUTO5 are interdependent with respect to their implementation in automated laboratory systems. The matrix describes the engineering relationships between the standards elements in each of the five documents. This matrix is provided so that designers and engineers, as well as users and customers, understand the relationships between the different standards' components in an automated system. The matrix format allows the users of one document to easily identify other standard elements, which relate to the standard elements in the document or documents from which they may be working, to design a system correctly.

How to Read the Matrix (See matrix on the next page.)

The numbers listed on the horizontal (X) and vertical (Y) axes contain multiple-digit numbers (e.g., (1)5.4, (5)5.4.1.3).

The 'first digit' (in parentheses) represents one of the five automation documents (e.g., (1)5.4 is from AUTO01; (5)5.4.1.3 is from AUTO5).

The 'remaining digits' represent the specific section of that document.

The symbol XX represents the direct 'engineering relationship' between two sections.

The symbol ## represents the section's 'self'; when it has been lined up with itself on the other axis.

SAMPLE

Related NCCLS Publications*

- AUTO2** **Laboratory Automation: Bar Codes for Specimen Container Identification.** This document provides specifications for use of linear bar codes on specimen tubes in the clinical laboratory and for use on laboratory automation systems.
- AUTO3** **Laboratory Automation: Communications with Automated Clinical Laboratory Systems, Instruments, Devices, and Information Systems.** The goal of this document is to facilitate accurate and timely electronic exchange of data and information between the automated laboratory elements.
- AUTO4** **Laboratory Automation: Systems Operational Requirements, Characteristics, and Information Elements.** This document provides standards of interest to operators for display of system status information such as specimen location, reagent supply, and warnings and alerts to support laboratory automation operations.
- AUTO5** **Laboratory Automation: Electromechanical Interfaces.** This document provides guidance for the standardization of electromechanical interfaces between instruments and/or specimen processing and handling devices and automation systems in the automated laboratory.
- GP2-A2-C** **NCCLS Procedure Manual Template.** This computer template enables laboratorians to prepare consistent technical procedures in the NCCLS format. The template and its user manual, used along with the GP2-A3 guideline, provide a procedure format that is as easy to use as a word processing program. Procedures can be stored as individual files for easy retrieval and updating, or they can be networked through the local computer system for electronic distribution throughout the laboratory. The template format consists of tables for recording essential information for all procedures and an outline of key headings for incorporating procedure-specific details.
- GP2-A3** **Clinical Laboratory Technical Procedure Manuals – Third Edition; Approved Guideline (1996).** This document provides guidance for the patient-testing community by addressing the design, preparation, maintenance, and use of paper or electronic technical procedure manuals.
- GP18-A** **Laboratory Design; Approved Guideline (1998).** This guideline provides a foundation of information about laboratory design elements that can be used to help define the issues being considered when designing a laboratory.
- GP19-A** **Laboratory Instruments and Data Management Systems: Design of Software User Interfaces and End-User Software Systems Validation, Operation, and Monitoring; Approved Guideline (1995).** The document identifies important factors that designers and laboratory managers should consider when developing new software-driven systems and selecting software user interfaces. Also included are simple rules to help prepare validation protocols for assessing the functionality and dependability of software.

* Proposed- and tentative-level documents are being advanced through the NCCLS consensus process; therefore, readers should refer to the most recent editions.

Related NCCLS Publications (Continued)

- H1-A4** **Evacuated Tubes and Additives for Blood Specimen Collection - Fourth Edition; Approved Standard (1996).** *American National Standard.* This standard discusses requirements for blood collection tubes and additives, including heparin, EDTA, and sodium citrate.
- H18-A2** **Procedures for the Handling and Processing of Blood Specimens; Approved Guideline—Second Edition (1999).** This guideline addresses multiple factors associated with handling and processing of specimens, and factors that can introduce imprecision or systematic bias into results.
- H38-P** **Calibration and Quality Control of Automated Hematology Analyzers; Proposed Standard (1999).** This document addresses calibration and quality control strategies for multichannel hematology analyzers; assignment of values to calibrator materials; calibration using stabilized blood controls; internal quality control; pair difference analysis; and use of the weighted moving average (\bar{x}_B) method.
- M29-A** **Protection of Laboratory Workers from Instrument Biohazards and Infectious Disease Transmitted by Blood, Body Fluids, and Tissue; Approved Guideline (1997).** A consolidation of M29-T2 and I17-P, this document provides guidance on the risk of transmission of hepatitis viruses and human immunodeficiency viruses in any laboratory setting; specific precautions for preventing the laboratory transmission of blood-borne infection from laboratory instruments and materials; and recommendations for the management of blood-borne exposure.

Explore the Latest Offerings From CLSI!

As we continue to set the global standard for quality in laboratory testing, we are adding products and programs to bring even more value to our members and customers.



By becoming a CLSI member, your laboratory will join 1,600+ other influential organizations all working together to further CLSI's efforts to improve health care outcomes. You can play an active role in raising global laboratory testing standards—in your laboratory, and around the world.

Find out which membership option is best for you at www.clsi.org/membership.



Find what your laboratory needs to succeed! CLSI U provides convenient, cost-effective continuing education and training resources to help you advance your professional development. We have a variety of easy-to-use, online educational resources that make eLearning stress-free and convenient for you and your staff.

See our current educational offerings at www.clsi.org/education.



When laboratory testing quality is critical, standards are needed and there is no time to waste. eCLIPSE™ Ultimate Access, our cloud-based online portal of the complete library of CLSI standards, makes it easy to quickly find the CLSI resources you need.

Learn more and purchase eCLIPSE at cli.org/eCLIPSE.

For more information, visit www.clsi.org today.