

3 December 2018

To: Recipients of EP14-A3

From: Jennifer K. Adams, MT(ASCP), MSHA
Vice President, Standards and Quality

Subject: Correction

This notification is to inform you of corrections made to CLSI document EP14-A3, *Evaluation of Commutability of Processed Samples, Approved Guideline—Third Edition*. The corrections are described below and shown as red text in the excerpt.

Appendix A, Description of Mathematical Model Used for Evaluating Commutability of Processed Samples Using Deming Regression:

Equation (7) was corrected to add a squared symbol after the first term under the radical and a hat above the sigma in the second term under the radical:

$$\hat{\beta}_H = \frac{\hat{\sigma}_Y^2 - \hat{\lambda}\hat{\sigma}_X^2 + \sqrt{(\hat{\sigma}_Y^2 - \hat{\lambda}\hat{\sigma}_X^2)^2 + 4\hat{\lambda}\hat{\sigma}_{XY}^2}}{2\hat{\sigma}_{XY}}$$

Equation (7) was also corrected in the following locations:

- Appendix C, beneath Figure C2 (p. 28)
- Appendix C, beneath Figure C5 (p. 35)
- Appendix C (p. 36)

Equation (7) also appears on p. 29; no corrections were needed in this instance.

If you require any additional clarification regarding these corrections, please contact CLSI Customer Service (customerservice@clsi.org).

We appreciate your commitment to CLSI and regret any inconvenience.